

# DICOM Correction Proposal

STATUS	Final Text
Date of Last Update	2019/03/23
Person Assigned	Wim Corbijn van Willenswaard
Submitter Name	Nikhilesh Sonar, <a href="mailto:nikhilesh.sonar@philips.com">nikhilesh.sonar@philips.com</a> Wim Corbijn van Willenswaard, <a href="mailto:wim.corbijn.van.willenswaard@philips.com">wim.corbijn.van.willenswaard@philips.com</a>
Submission Date	2016/01/11

Correction Number	CP-1552
Log Summary:	Resolve discrepancy in defined Value Constraint for ATNA message
Name of Standard	PS3.15 2019a + CP 1551
Rationale for Correction:	<p>The current definition for the <b>ParticipantObjectTypeCode</b> field –</p> <p style="text-align: center;"><b>Shall be: 2 = patient ID</b></p> <p>is non-compliant to the value type defined, which is CodedValueType. This also is inconsistent to the way the code is defined elsewhere in the section.</p> <p>There is no CID, or a triplet code, for Patient ID defined in the Standard currently.</p> <p>It would be prudent to rectify this inconsistency by defining the CID to meet the value type. In RFC-3881 this is defined as 2 with meaning Patient Number.</p> <p>It applies to the following tables:</p> <ul style="list-style-type: none"> <li>• Table A.5.3.3-1. Audit Message for Begin Transferring DICOM Instances</li> <li>• Table A.5.3.4-1. Audit Message for Data Export</li> <li>• Table A.5.3.5-1. Audit Message for Data Import</li> <li>• Table A.5.3.6-1. Audit Message for DICOM Instances Accessed</li> <li>• Table A.5.3.7-1. Audit Message for DICOM Instances Transferred</li> <li>• Table A.5.3.8-1. Audit Message for DICOM Study Deleted</li> </ul>
Correction Wording:	

*Item #1: Modify PS3.15 Table A.5.3.3-1, Table A.5.3.6-1, A.5.3.7-1, A-5.3.8-1. Change ParticipantObjectTypeCode description*

Participating Object: Patient (1)	ParticipantObjectTypeCode	M	Shall be: 1 = person
	ParticipantObjectTypeCodeRole	M	Shall be: 1 = patient

	ParticipantObjectDataLifeCycle	U	not specialized
	ParticipantObjectIDTypeCode	M	Shall be: <del>2 = patient ID</del> <u>EV (2, RFC-3881, "Patient Number")</u>
	ParticipantObjectSensitivity	U	not specialized
	ParticipantObjectID	M	The patient ID
	ParticipantObjectName	U	The patient name
	ParticipantObjectQuery	U	not specialized
	ParticipantObjectDetail	U	not specialized
	ParticipantObjectDescription	U	not specialized

*Item #2: Modify PS3.15 Table A.5.3.4-1, Table A.5.3.5-1. Change ParticipantObjectIDTypeCode description*

Participating Object: Patients (1..N)	ParticipantObjectTypeCode	M	Shall be: 1 = person
	ParticipantObjectTypeCodeRole	M	Shall be: 1 = patient
	ParticipantObjectDataLifeCycle	U	not specialized
	ParticipantObjectIDTypeCode	M	Shall be: <del>2 = patient ID</del> <u>EV (2, RFC-3881, "Patient Number")</u>
	ParticipantObjectSensitivity	U	not specialized
	ParticipantObjectID	M	The patient ID
	ParticipantObjectName	U	The patient name
	ParticipantObjectQuery	U	not specialized
	ParticipantObjectDetail	U	not specialized
	ParticipantObjectDescription	U	not specialized