

DICOM Correction Proposal

STATUS	Assigned
Date of Last Update	2019/05/12
Person Assigned	Kevin O'Donnell
Submitter Name	Kevin O'Donnell
Submission Date	2019/03/11

Correction Number	CP-1911
Log Summary:	Add Reason For Requested Procedure Query Keys to UPS
Name of Standard	PS3.4
Rationale for Correction:	IHE EBIW has highlighted a use case Reason for Requested Procedure query keys in UPS. They need to be changed from -/- to O/O in Table CC.2.5-3.
Correction Wording:	

Modify PS 3.4 C.2.5-3 as shown

Table CC.2.5-3. UPS SOP Class N-CREATE/N-SET/N-GET/C-FIND Attributes

Attribute Name	Tag	Req. Type N-CREATE (SCU/SCP)	Req. Type N-SET (SCU/SCP)	Final State	Req. Type N-GET (SCU/SCP)	Match Key Type	Return Key Type	Remark/Matching Type
Transaction UID	(0008,1195)	2/2 Shall be empty	(see CC.2.6.3)	O	Not allowed	-	-	Cannot be queried.
SOP Common Module								
...								
Unified Procedure Step Relationship Module								
...								
Referenced Request Sequence	(0040,A370)	2/2	Not allowed	O	3/2	R	2	Could be "changed" while SCHEDULED by canceling and re-creating with the "correct" values.
...								
>Reason for the Requested Procedure	(0040,1002)	3/3	3/3	O	-/3	<u>-O/O</u>	<u>-O/O</u>	

Attribute Name	Tag	Req. Type N-CREATE (SCU/SCP)	Req. Type N-SET (SCU/SCP)	Final State	Req. Type N-GET (SCU/SCP)	Match Key Type	Return Key Type	Remark/Matching Type
>-Reason for Requested Procedure Code Sequence	(0040,100A)	3/3	3/3	0	-/3	<u>-0/0</u>	<u>-0/0</u>	
>>Include Table CC.2.5-2a "UPS Code Sequence Macro"								
...								