

1	Status	Final Text
2	Date of Last Update	2020/01/16
3	Person Assigned	David Clunie
4		mailto:dclunie@dclunie.com
5	Submitter Name	David Clunie
6		mailto:dclunie@dclunie.com
7	Submission Date	2018/11/10

8	Correction Number CP-1928	
9	Log Summary: Add multi-energy CT attributes to de-identification profile	
10	Name of Standard	
11	PS3.15 2019e	
12	Rationale for Correction:	
13	Sup 188 did not address de-identification of the new Attributes that are descriptions, IDs and labels.	
14	Correction Wording:	

Add the following rows to DICOM PS3.15 Table E.1-1:

Table E.1-1. Application Level Confidentiality Profile Attributes

Attribute Name	Tag	Retd. (from ????)	In Std. Comp. IOD (from ????)	Basic Prof.	Rtn. Safe Priv. Opt.	Rtn. UIDs Opt.	Rtn. Dev. Id. Opt.	Rtn. Inst. Id. Opt.	Rtn. Pat. Chars. Opt.	Rtn. Long. Full Dates Opt.	Rtn. Long. Modif. Dates Opt.	Clean Desc. Opt.	Clean Struct. Cont. Opt.	Clean Graph. Opt.
Decomposition Description	(0018,937F)	N	Y	X								C		
Multi-energy Acquisition Description	(0018,937B)	N	Y	X								C		
Source End DateTime	(0018,936A)	N	Y	D						K	C			
Source Start DateTime	(0018,9369)	N	Y	D						K	C			
X-Ray Detector ID	(0018,9371)	N	Y	D			K							
X-Ray Detector Label	(0018,9373)	N	Y	X			K							
X-Ray Source ID	(0018,9367)	N	Y	D			K							