

## DICOM Correction Proposal

STATUS	Final Text
Date of Last Update	2020/01/17
Person Assigned	
Submitter Name	Uwe Tronnier, Sridhar Balasubramanian
Submission Date	2019/02/01

Correction Number	CP-1939
Log Summary: Semantic issue in consumable parts	
Name of Standard PS3.17 2019e	
Rationale for Correction: Node 1.18.1 denotes a “Contrast medium injection system manifold kit. Connecting patients with this kind of manifold kits always requires a special tube to the patient. This is normally included, when using a new kit. In case of using-on an existing kit, it would be necessary to use a new patient tube and therefore to add a 4 <sup>th</sup> consumable (SRT, A-26400, "Tube, device (physical object)") in the same way a new needle is used for each patient.  Content tree structure table: line of node 1.18.2.1 has code value (SRT, R-0038D, "No") in column Code Meaning or Example Value. At first, this is erroneous; hence R-0038D means “Yes”. “No” has Code R-00339. At second, it seems to be less work to change this “Yes” instead of introducing a 4 <sup>th</sup> consumable, which requires renumbering the rest of the example.	
Correction Wording:	

*Amend DICOM PS3.17 M444.1 as follows  
(changes to existing text are bold and underlined for additions and ~~struckthrough~~ for removals):*

Node	Code Meaning of Concept Name	Code Meaning or Example Value	Reference to TID/ CID/ Comments
...	...	...	...
1.18.2.1	Consumable is New	<del>(373067005</del> <u>373066001</u> , SCT, " <del>No</del> <u>Yes</u> ")	TID 11005 CID 231