

# DICOM Correction Proposal

STATUS	Letter Ballot
Date of Last Update	2020/01/16
Person Assigned	Clunie
Submitter Name	Silvia Winkler ( <a href="mailto:silvia.winkler@sigmasoft.at">silvia.winkler@sigmasoft.at</a> )
Submission Date	2019/06/26

Correction Number	CP-1973
Log Summary: There are different names used for the same UID in different parts of the standard. Same UIDs should be referred to by the same name.	
Name of Standard	PS3.4
<p>Rationale for Correction:</p> <p>Annex B in DICOM PS3.4 “Storage Service Class”  (<a href="http://dicom.nema.org/medical/dicom/current/output/html/part04.html#table_B.5-1">http://dicom.nema.org/medical/dicom/current/output/html/part04.html#table_B.5-1</a>) (normative)  SOP Class UID: 1.2.840.10008.5.1.4.1.1.88.59  SOP Class Name „<b>Key Object Selection Storage</b>”</p> <p>Annex A in DICOM PS3.6 „Registry of DICOM Unique Identifiers (UIDs)”  (<a href="http://dicom.nema.org/medical/dicom/current/output/html/part06.html#table_A-1">http://dicom.nema.org/medical/dicom/current/output/html/part06.html#table_A-1</a>) (normative).  UID Value: 1.2.840.10008.5.1.4.1.1.88.59  UID Name: „<b>Key Object Selection Document Storage</b>”</p> <p>IHE RAD-68 Table 4.68.4.1.2.3-1: XDS-I.b-specific Metadata Requirements  formatCode:  “This attribute shall be populated by the XDS-I Imaging Document Source as follows:  • Shall use “1.2.840.10008.5.1.4.1.1.88.59” (DICOM KOS SOP Class UID) as the  Format Code Value and “1.2.840.10008.2.6.1” (DICOM UID Registry UID) as the  Format Coding Scheme OID for a DICOM Manifest document.”</p> <p>The formatCode requires a displayName and it is not clear, what should be the correct displayName for the  Key Object Selection (Document) Storage SOP Class</p>	
Correction Wording:	

*change PS3.4 Table\_B.5-1*

...		
Key Object Selection <u>Document</u> Storage	1.2.840.10008.5.1.4.1.1.88.59	Key Object Selection Document IOD

