

DICOM Correction Proposal

STATUS	Letter Ballot
Date of Last Update	2020/01/16
Person Assigned	Tronnier
Submitter Name	Uwe Tronnier
Submission Date	2019/09/12

--

Correction Number	CP-1980
Log Summary: Change Requested Type attribute of row 15 within TID 11004 from Mandatory (M) to Optional (U)	
Name of Standard	PS3.16
Rationale for Correction: Within row 2, when one of the CID 12 values "Air", "Oxygen", "Water", "Carbon dioxide" is coded, one cannot give a value for row 15 (Unit of Presentation), because there is no presentation unit. The correct Requested Type should be U.	
Correction Wording: Change value of TID 11004, row 15, column "Req Type" from M to U.	

*Amend DICOM PS3.16 as follows
(changes to existing text are bold and underlined for additions and ~~struckthrough~~ for removals):*

TID 11004 Imaging Agent Component

Table TID 11004. Imaging Agent Component

	NL	Rel with Parent	VT	Concept Name	VM	Req Type	Condition	Value Set Constraint
15	>	CONTAINS	CODE	EV (732935002, SCT, "Unit of Presentation")	1	M <u>U</u>		DCID 68 "Imaging Agent Administration Pharmaceutical Unit of Presentation"