DICOM Correction Proposal

STATUS	Final Text
Date of Last Update	2024/03/24
Person Assigned	Kevin O'Donnell
Submitter Name	Kevin O'Donnell
Submission Date	2022/06/20

Correction Number	CP-2321

Log Summary: Add a code for Intraoperative Transducer to CID 12035

Name of Standard

PS3.16 2024a

Rationale for Correction:

CID 12035 contains codes to indicate the application for which a transducer is designed when describing a transducer.

Several vendors have transducers that are designed for intraoperative use. It is proposed to add a corresponding code.

Intraoperative probes are designed with strict temperature standards. Some intraoperative probes are inserted into the body through a laparoscope, typically inside the body cavity, and differ from insertion via catheter.

Identification and tracking of intraoperative probes can also be useful for infection control procedures.

Correction Wording:

Add code to CID 12035 Radiotherapeutic Dose Measurement Devices

CID 12035 Ultrasound Transducer Application

Version: 2019040920240324

Table CID 12035. Ultrasound Transducer Application

Coding Scheme Designator	Code Value	Code Meaning
DCM	125261	External Transducer
DCM	125262	Transesophageal Transducer
DCM	125263	Endovaginal Transducer
DCM	125264	Endorectal Transducer
DCM	125265	Intravascular Transducer
DCM	125266	Endocavitary Transducer
<u>DCM</u>	<u>125267</u>	Intraoperative Transducer

		_			_
Amend	PS	З.	16	Annex	' I)

Table D-1. DICOM Controlled Terminology Definitions

Code Value	Code Meaning	Definition	Notes
125261	External Transducer	Transducer is designed to be placed onto the surface of the subject.	
125262	Transesophageal Transducer	Transducer is designed for insertion into the esophagus.	
125263	Endovaginal Transducer	Transducer is designed for insertion into the vagina.	
125264	Endorectal Transducer	Transducer is designed for insertion into the rectum.	
125265	Intravascular Transducer	Transducer is designed for insertion into the subject via a catheter.	
125266	Endocavitary Transducer	Transducer is designed for insertion into either the vaginal cavity or the rectal cavity.	
<u>125267</u>	Intraoperative Transducer	Transducer is designed for insertion into the subject via an incision.	