

## DICOM Correction Proposal

STATUS	Final Text
Date of Last Update	2024-03-24
Person Assigned	Bjoern Nolte <a href="mailto:bjoern.nolte@siemens-healthineers.com">bjoern.nolte@siemens-healthineers.com</a>
Submitter Name	Ingo Wilkening <a href="mailto:ingo.wilkening@siemens-healthineers.com">ingo.wilkening@siemens-healthineers.com</a>
Submission Date	2022-07-19

Correction Number	CP-2326
Log Summary: Former existing coded-concept (F-10440,SRT,"Erect") in legacy data missing in PS3.16	
Name of Standard	PS3.16 2024a
<p>Rationale for Correction:</p> <p>Earlier editions of the DICOM standard contained in <a href="#">PS3.16, CID 19 Patient Orientation</a> a SNOMED RT Code of "F-10440" with the Code Meaning of "erect".</p> <p>This code was removed from the column "SNOMED-RT ID" when an NCIt Code was provided instead of a SNOMED CT code after it was inactivated by SNOMED.</p> <p>The codes (F-10440, SRT, "erect") and (F-10440, SNM3, "erect") may exist in legacy patient data stored in image archives (PACS).</p> <p>Current systems reading such legacy data will not be able to interpret that patient orientation of "F-10440" as it is not mentioned throughout the DICOM standard.</p> <p>As the SNOMED RT Code "F-10440" existed but is retired, it should be same added to <a href="#">DICOM PS3.16, Table J-1. SNOMED Codes Retired from DICOM Use</a>.</p>	

*Update PS3.16*

### Table J-1. SNOMED Codes Retired from DICOM Use

*Add new line for "F-10440":*

Retired Code Value	Code Meaning	Replacement Code	Notes
...			
<b><u>F-10440</u></b>	<b><u>erect</u></b>	<b><u>NCIt:C86043</u></b>	<b><u>Replacement code has meaning of "Upright".</u></b> <b><u>Retired code is inactive in SNOMED CT.</u></b>
...			