

# DICOM Correction Proposal

STATUS	May 2024 Voting Packet
Date of Last Update	2024/03/20
Person Assigned	<a href="mailto:Bjoern.nolte@siemens-healthineers.com">Bjoern.nolte@siemens-healthineers.com</a>
Submitter Name	Antje Schroeder ( <a href="mailto:aschroeder@siemens-healthineers.com">aschroeder@siemens-healthineers.com</a> )
Submission Date	2023/05/31

Correction Number	CP-2315
Log Summary:	Clarify Single Conformance Statement in PS3.2
Name of Standard	DICOM PS3.2 2023b - Conformance
Rationale for Correction:	Current PS3.2 states in Section 6 of PS3.2 "A single Conformance Statement shall be provided". A clarification is needed that this this means that the conformance statement doesn't need to be in a single file and can be distributed across multiple files.
Correction Wording:	<include proposed change below, following guidelines for formatting of changes in supplements>

*Modify Section 6 of PS3.2 as indicated below*

The content of Conformance Statement uses a consistent structure regardless of whether the implementation supports a DIMSE interface, a DICOM Media Storage interface, a DICOM Web Service interface, DICOM Real Time Video interface, or a combination thereof. A single Conformance Statement shall be provided with the appropriate sections filled in. Sections not relevant for the implementation shall be kept and marked as not applicable. Subsections of a section marked as not applicable need not be included in the Conformance Statement (see the template in Annex N).

## **Note**

**A conformance statement is a single document or a main document with possibly a number of annexes with appropriate references.**

The first part of the Conformance Statement contains a DICOM Conformance Statement Overview ...