

1	Status	Letter Ballot
2	Date of Last Update	2025/03/30
3	Person Assigned	David Clunie
4		mailto:dclunie@dclunie.com
5	Submitter Name	David Clunie
6		mailto:dclunie@dclunie.com
7	Submission Date	2024/10/23

8	Correction Number CP-2459	
9	Log Summary: More data elements to de-identify	
10	Name of Standard	
11	PS3.15	
12	Rationale for Correction:	
13	Some device related identifiers are not included (and should be retained only with retain device identifiers option:	
14	(0018,1010) Secondary Capture Device ID	
15	(0018,1011) Hardcopy Creation Device ID	
16	Correction Wording:	

Amend DICOM PS3.15 as follows (changes to existing text are bold and underlined for additions and ~~struckthrough~~ for removals):

### E.1.1 De-identifier

**Table E.1-1. Application Level Confidentiality Profile Attributes**

Attribute Name	Tag	Retd. (from PS3.6)	In Std. Comp. IOD (from PS3.3)	Basic Prof.	Rtn. Safe Priv. Opt.	Rtn. UIDs Opt.	Rtn. Dev. Id. Opt.	Rtn. Inst. Id. Opt.	Rtn. Pat. Chars. Opt.	Rtn. Long. Full Dates Opt.	Rtn. Long. Modif. Dates Opt.	Clean Desc. Opt.	Clean Struct. Cont. Opt.	Clean Graph. Opt.
Device Alternate Identifier	(3010,001B)	N	Y	Z										
Device Description	(0050,0020)	N	Y	X			K							
Device Label	(3010,002D)	N	Y	D			K							
Device Serial Number	(0018,1000)	N	Y	X/Z/D			K							
Device Setting Description	(0016,004B)	N	Y	X								C		
Device UID	(0018,1002)	N	Y	U		K	K							
<b><u>Hardcopy Creation Device ID</u></b>	<b><u>(0018,1011)</u></b>	<b><u>Y</u></b>	<b><u>N</u></b>	<b><u>X</u></b>			<b><u>K</u></b>							
Manufacturer's Device Class UID	(0018,100B)	N	Y	U		K	K							
Manufacturer's Device Identifier	(3010,0043)	N	Y	Z			K							
<b><u>Secondary Capture Device ID</u></b>	<b><u>(0018,1010)</u></b>	<b><u>N</u></b>	<b><u>Y</u></b>	<b><u>X</u></b>			<b><u>K</u></b>							

For reference unchanged DICOM PS3.3:

### C.8.6.1 SC Equipment Module

**Table C.8-24. SC Equipment Module Attributes**

Attribute Name	Tag	Type	Attribute Description
...			
Secondary Capture Device ID	(0018,1010)	3	User defined identification of the device that converted the image.
Secondary Capture Device Manufacturer	(0018,1016)	3	Manufacturer of the Secondary Capture Device.
Secondary Capture Device Manufacturer's Model Name	(0018,1018)	3	Manufacturer's model number of the Secondary Capture Device.
Secondary Capture Device Software Versions	(0018,1019)	3	Manufacturer's designation of software version of the Secondary Capture Device.
...			

For reference unchanged DICOM PS3.6:

## 6 Registry of DICOM Data Elements

Table 6-1. Registry of DICOM Data Elements

Tag	Name	Keyword	VR	VM	
(0018,1011)	Hardcopy Creation Device ID	HardcopyCreationDeviceID	LO	1	RET (2007)