

# DICOM Correction Proposal

STATUS	Letter Ballot
Date of Last Update	2026/01/28
Person Assigned	Clunie
Submitter Name	Van Wijk
Submission Date	2022/03/31

Correction Number	CP-2253
Log Summary: Clarify description of Image Type and Frame Type in Enhanced MR Image Module	
Name of Standard	
PS3.3	
Rationale for Correction:	
For Enhanced MR Image Module, Annex C.8.13.1.1.1 Image Type and Frame Type, the standard states:	
Image Type (0008,0008) and Frame Type (0008,9007) are not included in this Macro but one or the other is always included in the Module or Macro that invokes this Macro, and they are therefore described here.	
The text makes it appear as if this concerns a macro, where the actual context is the Enhanced MR Image Module (as per Annex C.8.13.1 – Table C.8-79. Enhanced MR Image module Attributes).	
Correction Wording:	

Amend PS3.3:

## C.8.13.1.1 Enhanced MR Image Module Attribute Description

### C.8.13.1.1.1 Image Type and Frame Type

Image Type (0008,0008) is included in this Module, but and Frame Type (0008,9007) are is not included in this Macro, but one or the other is always included in the Module or Macro that invokes this Macro, and they both are therefore described here.