


# DICOM Correction Proposal

STATUS	Letter Ballot
Date of Last Update	2026/01/28
Person Assigned	Kevin O'Donnell
Submitter Name	Kevin O'Donnell
Submission Date	2025/04/22

Correction Number	CP-2556												
Log Summary: Add Placenta to Cervix Internal Os Distance Measurement													
Name of Standard													
PS 3.16													
Rationale for Correction:													
<p>Some Ultrasound OB/Gyn exams include a measurement of the distance from the leading placental edge to the internal cervical os (as part of assessment for low-lying placenta).</p> <p><a href="https://www.researchgate.net/publication/319869471_Measuring_leading_placental_edge_to_internal_cervical_os_Transabdominal_versus_transvaginal_approach">https://www.researchgate.net/publication/319869471_Measuring_leading_placental_edge_to_internal_cervical_os_Transabdominal_versus_transvaginal_approach</a></p> <p>A corresponding code is added to CID 12011 Ultrasound Pelvis and Uterus Measurement. CID 12011 is included in Row 3 of TID 5015 (Pelvis and Uterus Section) as measurement codes for TID 300.</p>													
	<table border="1"> <tr> <td>①</td> <td>Cervix</td> </tr> <tr> <td>②</td> <td>External Os</td> </tr> <tr> <td>③</td> <td>Internal Os</td> </tr> <tr> <td>④</td> <td>Bladder</td> </tr> <tr> <td>⑤</td> <td>Amniotic fluid</td> </tr> <tr> <td>⑥</td> <td>Placenta</td> </tr> </table>	①	Cervix	②	External Os	③	Internal Os	④	Bladder	⑤	Amniotic fluid	⑥	Placenta
①	Cervix												
②	External Os												
③	Internal Os												
④	Bladder												
⑤	Amniotic fluid												
⑥	Placenta												
Correction Wording:													

Add code to CID 12011 as shown

## CID 12011 Ultrasound Pelvis and Uterus Measurement

**Keyword:** UltrasoundPelvisAndUterusMeasurement  
**FHIR Keyword:** dicom-cid-12011-UltrasoundPelvisAndUterusMeasurement  
**Type:** Extensible  
**Version:** 20230130yyyymmdd  
**UID:** 1.2.840.10008.6.1.559

Table CID 12011. Ultrasound Pelvis and Uterus Measurement

Coding Scheme Designator	Code Value	Code Meaning	UMLS Concept Unique ID
LN	11961-0	Cervix Length	C0552078
LN	12145-9	Endometrium Thickness	C0552258

Coding Scheme Designator	Code Value	Code Meaning	UMLS Concept Unique ID
DCM	<u>Newcode1</u>	<u>Internal Os Distance</u>	

*Add definitions to PS 3.16 Annex D*

**Table D-1. DICOM Controlled Terminology Definitions**

Code Value	Code Meaning	Definition	Notes
...			
<u>Newcode1</u>	<u>Internal Os Distance</u>	<u>The distance from the leading edge of the placenta to the internal os of the cervix.</u> <u>Sometimes abbreviated IOD.</u>	